





# EU Type Examination Certificate

This is to certify that:

Sunbeam International GmbH Schumanstr. 12 Würselen 52146 Germany

Holds Certificate Number:

CE 730303

In respect of:

#### Model HYGISUN HS0501A Face mask To technical specification Annex II (EHSR) of the PPE Regulation (EU) 2016/425 PPE for use by healthcare professionals as per Commission recommendation 2020/403

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2020-07-03 Latest Issue: 2020-07-03

Drs. Dave Hagenaars, Managing Director

Effective Date: 2020-07-03 Expiry Date: 2021-07-03

Page: 1 of 3



# ...making excellence a habit."

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated <u>online</u>.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands A member of BSI Group of Companies.

# EU Type Examination Certificate

No. CE 730303

## **Product Specification**

Product Name:	Particulate Respirator.	
Product Type:	Particulate filtering half masks for use by Healthcare professionals.	
Model:	HYGISUN HS0501A.	
Classification:	FFP2 NR un-valved.	
Technical Specification:	Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.	
Product Description:	The respirator is non-reusable, secured to the face of the user by a pair of elasticated ear straps, and has no exhalation valve. The respirator is FFP2 class, vertical fold flat type.	
	The respirator listed on this certificate is for use by healthcare workers, first responders and other personnel involved in the efforts to contain the COVID-19 virus and avoid its further spread.	
	The product covered by this certificate is not approved for industrial applications and the certificate is only valid as long as EU Commission recommendation sheet 2020/403 remains applicable.	
Product Assessments:	BSI's PPE for Healthcare Professionals 2020/403 – RPE Technical Specification.	

First Issued: 2020-07-03 Latest Issue: 2020-07-03 Effective Date: 2020-07-03 Expiry Date: 2021-07-03

Page: 2 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated <u>online</u>.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands A member of BSI Group of Companies.

# EU Type Examination Certificate

No. CE 730303

### **Certificate Administration Details**

Technical File Reference: Sunbeam International GmbH, TCF.01, V0 dated 28/06/2020.

### **Certificate Amendment Record:**

Issue date	Comments	BSI Review No.
July 2020	First issue.	2797:20:3220783

### **Certificate validity**

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module C2) as referenced on BSI issued Certificate CE 730304.

First Issued: 2020-07-03 Latest Issue: 2020-07-03 Effective Date: 2020-07-03 Expiry Date: 2021-07-03

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands A member of BSI Group of Companies.